

The National Marine Biological  
Analytical Quality Control Scheme  
[www.nmbaqcs.org](http://www.nmbaqcs.org)

Particle Size Results – PS32

David Hall  
Unicomarine Ltd.  
October 2008  
E-mail: [davidhall@unicomarine.com](mailto:davidhall@unicomarine.com)

  
*unicomarine*

## **Contents**

### **Tables**

- Table 1. Summary of the particle size information received from participating laboratories and replicate analysis laboratories for the thirty-second particle size distribution – PS32.
- Table 2. Z-score results for the derived statistics supplied by participating laboratories for the particle size exercise – PS32.

### **Figures**

- Figure 1. Particle size distribution curves resulting from analysis of fourteen replicate samples distributed as PS32.
- Figure 2. Particle size distribution curves from participating laboratories for sediment samples from PS32.
- Figure 3. Z-scores for PS32 derived statistics (replicated data not displayed).

**Table 1. Summary of the particle size information received from participating laboratories and replicate analysis laboratories for the thirty-second particle size distribution - PS32.**

**Replicate Sample Data**

Sample	Method	% < 63 micron	Median (phi)	Mean (phi)	Sort	IGS (SKi)
PS32_42	L <sup>1</sup>	58.78	4.88	4.79	2.62	-0.02
PS32_43	L <sup>1</sup>	57.88	4.82	4.75	2.63	0.00
PS32_44	L <sup>1</sup>	59.88	4.97	4.85	2.62	-0.04
PS32_45	L <sup>1</sup>	58.96	4.91	4.81	2.62	-0.02
PS32_46	L <sup>1</sup>	60.02	4.99	4.86	2.61	-0.04
PS32_47	L <sup>1</sup>	59.29	4.94	4.82	2.62	-0.03
PS32_48	L <sup>1</sup>	57.23	4.76	4.72	2.64	0.01
PS32_35	L <sup>2</sup>	100	7.08	7.204	1.302	0.215
PS32_36	L <sup>2</sup>	100	7.15	7.279	1.289	0.203
PS32_37	L <sup>2</sup>	100	7.45	7.593	1.151	0.247
PS32_38	L <sup>2</sup>	100	7.15	7.269	1.292	0.201
PS32_39	L <sup>2</sup>	100	7.43	7.582	1.149	0.254
PS32_40	L <sup>2</sup>	100	7.41	7.564	1.144	0.263
PS32_41	L <sup>2</sup>	100	7.10	7.221	1.305	0.206
UM	Overall Average	79.43	6.07	6.09	1.93	0.10
UM	L <sup>1</sup> RepAv	58.86	4.90	4.80	2.62	-0.02
UM	L <sup>2</sup> RepAv	100	7.25	7.39	1.23	0.23

**Participant Data**

Lab	Method	%<63μm	Median	Mean	Sort	IGS (SKi)
LB1501	L	62.02	4.98	4.69	2.54	-2.4
LB1502	-	-	-	-	-	-
LB1503	L	68.82	5.56	5.33	2.52	-0.09
LB1504	L	85.50	5.30	5.24	1.93	-0.1
LB1505	L	77.83	6.17	5.97	2.35	0.1
LB1506*	L	67.16	5.56	5.25	2.48	-0.16
LB1507	L	67.16	5.56	5.25	2.48	-0.16
LB1508	L	70.54	5.97	5.51	2.49	-0.28
LB1509	L	89.61	6.318	8.926	1.925	0.06
LB1510	L	72.10	5.86	5.50	2.45	-0.18
LB1511	L	69.69	5.64	5.44	2.59	-0.09
LB1526	L	65.53	5.34	3.49	2.53	0.06
LB1527	L	85.355	6.07	5.99	1.84	-0.07

Key to methods:

L - Laser analysis	DS - Dry sieve	CC - Coulter counter
S - Sieve	WS - Wet sieve	FD - Freeze dried
P - Pipette		

L<sup>1</sup> - Replicate analysis by Malvern Laser (MS2000)

L<sup>2</sup> - Replicate analysis by Coulter Laser (LS230)

L\* - data for this laboratory not included in calculations (replicate data).

"-" - No data. See forthcoming Annual Report, Section 6, for details.

Shaded cells - maximum and minimum values for each derived statistic.

**Table 2. Z-score results for the derived statistics supplied by participating laboratories for the particle size exercise - PS32.**

PS32	LabCode	%<63µm	z-score	Flag	Median	z-score	Flag	Mean	z-score	Flag	Sort	z-score	Flag	IGS (SKi)	z-score	Flag	Description: pre/post analysis
Laser <sup>1</sup> RepAv		58.86	-1.65	PASS	4.90	-1.88	PASS	4.80	-1.04	PASS	2.62	1.01	PASS	-0.019	0.45	PASS	Sandy mud / Sandy mud
Laser <sup>2</sup> RepAv		100	2.83	Fail	7.25	3.64	Fail	7.39	2.52	Fail	1.23	-3.86	Fail	0.23	2.62	Fail	Mud / Mud
LB1501		62.02	-1.31	PASS	4.98	-1.69	PASS	4.69	-1.19	PASS	2.54	0.73	PASS	-2.4	-20.33	Fail	Drk Brwn, consolidated, slightly sticky, very slightly gravelly, sandy mud, with organic fragments and mica present(g)sM
LB1502		-	-	Deemed Fail	-	-	Deemed Fail	-	-	Deemed Fail	-	-	Deemed Fail	-	-	Deemed Fail	- / -
LB1503		68.82	-0.57	PASS	5.56	-0.32	PASS	5.33	-0.31	PASS	2.52	0.66	PASS	-0.090	-0.18	PASS	sM / sM
LB1504		85.50	1.25	PASS	5.30	-0.93	PASS	5.24	-0.43	PASS	1.93	-1.41	PASS	-0.100	-0.26	PASS	Mud / Mud
LB1505		77.83	0.42	PASS	6.17	1.11	PASS	5.97	0.57	PASS	2.35	0.07	PASS	0.100	1.48	PASS	Sandy mud (anoxic) / Sandy mud (sM)
LB1506*		67.16	-0.75	PASS	5.56	-0.32	PASS	5.25	-0.42	PASS	2.48	0.52	PASS	-0.16	-0.79	PASS	Mud / Sandy Mud
LB1507		67.16	-0.75	PASS	5.56	-0.32	PASS	5.25	-0.42	PASS	2.48	0.52	PASS	-0.160	-0.79	PASS	Mud / Sandy Mud
LB1508		70.54	-0.38	PASS	5.97	0.64	PASS	5.51	-0.06	PASS	2.49	0.56	PASS	-0.280	-1.83	PASS	sandy mud / sandy mud
LB1509		89.61	1.70	PASS	6.32	1.45	PASS	8.926	4.63	Fail	1.93	-1.42	PASS	0.060	1.13	PASS	Silt / Poorly sorted medium silt
LB1510		72.10	-0.21	PASS	5.86	0.38	PASS	5.50	-0.08	PASS	2.45	0.42	PASS	-0.180	-0.96	PASS	Sandy mud / Sandy mud
LB1511		69.69	-0.47	PASS	5.64	-0.14	PASS	5.44	-0.16	PASS	2.59	0.91	PASS	-0.090	-0.18	PASS	Sandy MUD / Sandy MUD
LB1526		65.53	-0.92	PASS	5.34	-0.84	PASS	3.49	-2.84	Fail	2.53	0.70	PASS	0.060	1.13	PASS	Silt / -
LB1527		85.36	1.24	PASS	6.07	0.87	PASS	5.99	0.60	PASS	1.84	-1.72	PASS	-0.070	0.00	PASS	Mud / Sandy Mud


"-" no return and/or data from laboratory. See text, Section 6 in the forthcoming Annual Report, for details.

"\*" = centralised analysis


Laser<sup>1</sup>RepAv - Average of replicate analysis by Malvern Laser (Mastersizer2000)

Laser<sup>2</sup>RepAv- Average of replicate analysis by Coulter Laser (LS230)

**Figure 1. Particle size distribution curves resulting from analysis of fourteen replicate samples of sediment distributed as PS32. Seven samples analysed by Malvern Laser and seven samples analysed by Coulter Laser.**



**Figure 2. Particle size distribution curves from participating laboratories for sediment samples from PS32 (replicated data excluded).**



**Figure 3. Z-scores for PS32 derived statistics (replicated data not displayed).**

